

Certificate of Analysis

TEST DATE: March 21, 2011

PRODUCT: 100 mOsm/kg H₂O
Calibration Standard

PART NUMBER: 3LA011

LOT #: 0304021

EXPIRATION DATE: 2014-03

Product Data

Test Method	Analytical Result	Specification
Osmolality	100.5 mOsm /kg H ₂ O	100 ± 2.0 mOsm/kg H ₂ O

The Quality Management System of Advanced Instruments, Inc. and its subsidiaries are certified according to the requirements of ISO 9001 and ISO 13485. The product defined herein has been manufactured in accordance with Quality System requirements and conforms to all specifications.

This solution is traceable to standards prepared from National Institute of Standards and Technology (NIST) Standard Reference Material (SRM) 919b Sodium Chloride.

Balances used to prepare these standards have been calibrated with American Society for Testing and Materials (ASTM) Class weights traceable to NIST.



Andrea Curria

Quality Control