



Certificate of Analysis

Test Date: 9 December 2015
 Product: 200 mOsm/kg Calibration Standard
 Part Number: 3MA020
 Lot Number: 0515C294
 Expiration Date: 2018-11

Product Data

Test Method	Analytical Result	Specification
Osmolality (Freezing Point Depression)	200 mOsm/kg H ₂ O	200 ± 2.0 mOsm/kg H ₂ O

The Quality Management System of Advanced Instruments, Inc. and its subsidiaries is certified according to the requirements of ISO 9001 and ISO 13485. The product defined herein has been manufactured in accordance with Quality System requirements and conforms to all specifications.

Analyst	Approver
Printed Name: <u>Mark Hunter</u>	Printed Name: <u>MEREDITH PESTA</u>
Signature: <u></u>	Signature: <u></u>

Traceability

This product is traceable to NIST (SRM) 919b Sodium Chloride via an unbroken chain of comparisons. The NIST standards were formulated gravimetrically following USP Chapter 785 and tested on a Quality Control Freezing Point Depression Osmometer.

NIST Traceable Standard Used	Standard Lot Number	Standard Expiration
290 mOsm/kg H ₂ O	0701215	2016-07

Balances used to prepare these NIST standards have been calibrated with American Society for Testing and Materials (ASTM) class weights traceable to NIST.

Note: NIST standard formulation and balance calibration records are on file.

Hazardous Information: Refer to the Safety Data Sheet (SDS).

Statement of Uncertainty: Refer to document DCN11-1283.

Storage and Handling: The information on this Certificate is only valid when the manufacturer's instructions are followed. Please see the product insert for more information.