

### Certificate of Analysis

Test Date:	14 November 2014		
Product:	Fluorophos® Alkaline Phosphatase (ALP) Reagent Set		
Part Number:	FLA224/FLA225		
Kit Lot Number:	0904C017		
Buffer Lot Number:	1184C270		
Substrate Lot Number:	1114C500		
Expiration Date:	2016-09		
<b>Product Data</b>			
<u>Test Method</u>	<u>Analytical Result</u>	<u>Specification</u>	
Alkaline Phosphatase (ALP) Reagent Set - A/D Reading	614 FLU	< 1200 FLU's	
Alkaline Phosphatase (ALP) Reagent Set Fluorometer Measurement	<10 mU/L	< 10 mU/L	
The Quality Management Systems of Advanced Instruments, Inc. is registered as conforming to the requirements of ISO 9001 and ISO 13485.			
Analyst	Approver		
Printed Name: <u>Mark Hunter</u>	Printed Name: <u>MEREDITH PESTA</u>		
Signature: <u><i>Mark Hunter</i></u>	Signature: <u><i>Meredith Pesta</i></u>		

**Storage and Handling:** The information on this Certificate is only valid when the manufacturer's instructions for the product are followed. Please see the product insert for more information.

**Hazardous Information:** Refer to the Material Safety Data Sheet (MSDS).