


Certificate of Analysis


Test Date: 7 October 2014
Product: Renol™ 800 mOsm/kg H₂O Urine Osmolality Controls
Part Number: 200219
Lot Number: 0934C292
Expiration Date: 2016-09

Product Data

Test Method	Analytical Result	Specification
800 Osmolality (Freezing Point Depression)	799 mOsm/kg H ₂ O	800 ± 10.0 mOsm/kg H ₂ O

The Quality Management System of Advanced Instruments, Inc. and its subsidiaries is certified according to the requirements of ISO 9001 and ISO 13485. The product defined herein has been manufactured in accordance with Quality System requirements and conforms to all specifications.

Analyst
Printed Name: Mark Hunter
Signature: 

Approver
Printed Name: MEREDITH PESTA
Signature: 

Hazardous Information: Refer to the Material Safety Data Sheet (MSDS).

Storage and Handling: The information on this Certificate is only valid when the manufacturer's instructions are followed. Please see the product insert for more information.