

Certificate of Analysis

TEST DATE: November 3, 2010

PRODUCT: 422 m°H (-408 m°C)
Calibration Standard

PART NUMBER: 3LA022

LOT #: 1005050

EXPIRATION DATE: 2012-10

Product Data


Parameter	Analytical Result	Specification
Freezing Point Depression	422.8 m°H	422 ± 2.0 m°H

The Quality Management System of Advanced Instruments, Inc. and its subsidiaries is certified according to the requirements of ISO 9001:2000 and ISO 13485:2003. The product defined herein has been manufactured in accordance with Quality System requirements and conforms to all specifications.

This solution is traceable to standards prepared from National Institute of Standards and Technology (NIST) Standard Reference Material (SRM) 919b Sodium Chloride.

Balances used to prepare these standards have been calibrated with American Society for Testing and Materials (ASTM) Class weights traceable to NIST.

This product is prepared in accordance with AOAC Official Method 961.07.



Meredith E. Pesta
Quality Control