

Certificate of Analysis

Test Date: 12 June 2015
Product: Protinol® 280 mOsm/kg H₂O Protein-Based Osmolality Control
Part Number: 200215
Lot Number: 2395C287
Expiration Date: 2016-11

Product Data

Test Method	Analytical Result	Specification
280 Osmolality (Freezing Point Depression)	276 mOsm/kg H ₂ O	280 ± 7.0 mOsm/kg H ₂ O

The Quality Management System of Advanced Instruments, Inc. and its subsidiaries is certified according to the requirements of ISO 9001 and ISO 13485. The product defined herein has been manufactured in accordance with Quality System requirements and conforms to all specifications.

Analyst
Printed Name: MEREDITH PESTA Approver
Printed Name: Mark Hunter
Signature: Meredith Pesta Signature: Mark Hunter

Hazardous Information: Refer to the Material Safety Data Sheet (MSDS).

Storage and Handling: The information on this Certificate is only valid when the manufacturer's instructions are followed. Please see the product insert for more information.